

# **AC DIGITAL CLAMP METER KEW 2200R / 2200**



#### KEW 2200R / 2200 Specifications

	KEW 2200R	KEW 2200
Detection method	RMS	Averaging value
AC A	40.00/400.0/1000A(Auto-ranging) ±1.5%rdg±5dgt(45~65Hz) ±2.0%rdg±5dgt(40Hz~1kHz)	40.00/400.0/1000A(Auto-ranging) ±1.4%rdg±6dgt(50/60Hz) ±1.6%rdg±6dgt(45~65Hz)
AC V	4.000/40.00/400.0/600V(Auto-ranging) ±1.8%rdg±7dgt(45~65Hz) ±2.3%rdg±8dgt(65~500Hz)	
DC V	400.0mV/4.000/40.00/400.0/600V(Auto-ranging) ±1.0%rdg±3dgt* *400mV range is excluded	
Ω	$400.0\Omega/4.000/40.00/400.0k\Omega/4.000/40.00M\Omega$ (Auto-ranging) $\pm 2.0$ %rdg $\pm 4$ dgt(0~400kΩ) $\pm 4.0$ %rdg $\pm 4$ dgt(4MΩ) $\pm 8.0$ %rdg $\pm 4$ dgt(40MΩ)	
Continuity buzzer	buzzer sounds below $50\pm30\Omega$	
Conductor size	φ33mm max.	
Applicable standards	IEC61010-1 CATIV300V', CATII600V Pollution degree2(AC A) *2200R only CATII300V, CATII600V Pollution degree2(AC/DC V) IEC61010-031, IEC61010-2-032, IEC61326(EMC), EN50581(RoHS)	
Power source	R03/LR03(AAA)(1.5V)×2	
Continuous measuring time	Approx.120 hours	Approx.350 hours
	Auto power off: approx.10 minutes	
Dimensions/Weight	190(L)x68(W)x20(D)mm / Approx.120g(including batteries)	
Accessories	7107A (Test leads), 9160 (Carrying case), R03(AAA)×2, Instruction manual	
Optional	8008 (Multi-tran)	

#### Reliable result can be obtained even from distorted wave form generated by nowadays electronic loads True RMS type Averaging value type Waveform √correct √correct reading reading √correct higher reading reading √correct lower reading reading √correct lower reading reading

#### Accessories






### Optional











Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely Safety Warnings in the instrument thoroughly safety warnings in th to operate the instrument on a correct power supply and voltage rating marked on each instrument.

# COSINUS Messtechnik - Ihr Partner für Messlösung in allen elektrischen und physikalischen Anwendungen

## **COSINUS Messtechnik GmbH**

Rotwandweg 4 82024 Taufkirchen

Tel.: 089 / 66 55 94 - 0 Fax: 089 / 66 55 94 - 30